

Nederlands Meetinstituut

**Member State**  
The Netherlands

**OIML Certificate N°R60/1991-NL-97.24**  
Project number 10077107  
Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

**Issuing authority**

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1, Dordrecht  
Person responsible: A.J. Nederlof

**Applicant**

Name: Revere Transducers  
Address: Ramshoorn 7  
4824 AG Breda  
The Netherlands

**Manufacturer of the certified pattern**

Name: Revere Transducers  
Address: Ramshoorn 7  
4824 AG Breda  
The Netherlands

**Identification of the certified pattern**

Type : SSB  
 $P_i = 0.7$

Maximum Capacity ( $E_{max}$ )	500, 1000, 2000, 5000 and 10000 kg				2000, 5000 and 10000 kg			
Accuracy Class	C							
Maximum number of load cell intervals (n)	1000	2000	3000	3000 (MR)	3000	3000	4000	4000 (MR)
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	10000		20000		15000		10000	20000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	--		--		6000	8000	--	--

**Nederlands Meetinstituut**  
**Hugo de Grootplein 1**  
**3314 EG Dordrecht (NL)**  
Telephone +31 78 6332332  
Telefax +31 78 6332309

NMI B.V. (Chamber of Commerce  
Haaglanden No. 27228701)


Subsidiary companies:  
NMI Certin B.V. (27233418)  
NMI Van Swinden Laboratorium B.V. (27228703)  
NMI Inspecties en Kansspeltechniek B.V. (27228700)  
NMI International B.V. (27239176)

This certificate is issued under the provision that  
NMI B.V. nor its subsidiary companies accept any  
liability.

Reproduction of the complete certificate is allowed.  
Parts of the certificate may only be reproduced  
after written permission.



**QUALIFIED  
BY STERLAB**  
Reg. nr. L 029



Nederlands Meetinstituut

**Member State**  
The Netherlands

**OIML Certificate N°R60/1991-NL-97.24**

Project number 10077107

Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC2663 and the appertaining documentation folder, with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60  
edition 1991 (E)  
for accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the Loadcell concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report:

- N° R60/1991-NL-95.13, that includes 38 pages;
- N° R60/1991-NL-97.24A, that includes 35 pages;
- N° R60/1991-NL-97.24B, that includes 36 pages.

The issuing authority  
A.J. Nederlof

*KS*

12 August 1997

The OIML member  
G.J. Faber

12 August 1997

\*  
\* \*

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.